Chapter 837

Deming Regression

Introduction

This procedure provides sample size and power calculations for (unweighted) Deming regression when the goal of the study is to test whether the intercept is zero and the slope is one. An additional, simplifying assumption is made that the measurement standard deviation is constant throughout the range of the data. These results are based on the work of Linnet (1999).

Technical Details

Deming Regression

Deming regression is a technique for fitting a straight line to two-dimensional data where both variables, *X* and *Y*, are measured with error. This is different from simple linear regression where only the response variable, *Y*, is measured with error. Deming regression is often used for method comparison studies in clinical chemistry to look for systematic differences between two measurement methods. It is assumed that *X* represents the existing method and *Y* represents the new method.

Deming regression uses paired measurements, (x_i, y_i) , measured with errors, ε_i and δ_i , where

$$x_i = X_i + \varepsilon_i$$
$$v_i = Y_i + \delta_i.$$

to estimate the intercept, A, and the slope, B, in the equation

$$\hat{Y}_i = A + B\hat{X}_i.$$

 $\hat{X_i}$ and $\hat{Y_i}$ are estimates of the "true" (or expected) values of X_i and Y_i , respectively.

This procedure assumes that the measurement error ratio, $\lambda = V(\varepsilon_i)/V(\delta_i)$, is equal to one. Because of this, it gives the same result as orthogonal regression. It provides results using the unweighted method.

Hypotheses Concerning the Systematic Differences Between Methods

Using the Deming regression equation, the systematic difference, D_c , between the two methods at a particular level X_c is given by

$$D_c = Y_{est_c} - X_c = A + (B - 1)X_c$$

 D_c is zero if A = 0 and B – 1 = 0. These two hypotheses can be tested using two t-tests:

$$t_A = (\hat{A} - 0)/SE(\hat{A})$$
$$t_B = (\hat{B} - 1)/SE(\hat{B})$$

One task that must be completed during the planning phase is to determine the size of D_c that is medically significant. This value can then be used to determine the corresponding values of A and B that should be used in the power calculations.

Experimental Design

This procedure assumes a very specific experimental design in which *N* objects are measured using both an existing method, *X*, and a new method, *Y*. To provide an estimate of the measurement error, each measurement is duplicated. Hence, each object is actually measured four times, twice using method *X* and twice using method *Y*.

Power of the Test that B = 1

This section will provide the formulas for calculating the power for the tests of B = 1 and A = 0. These results come from Linnet (1999, page 893).

The power of this test is computed as follows. First, compute t_B using

$$t_{B} = \frac{B-1}{CVc_{B}} \sqrt{\frac{N}{2}} - |t_{\alpha/2,df}|$$
$$CV = \frac{\sigma_{X}}{x_{m}}$$
$$c_{B} = \sqrt{\frac{x_{m}^{2}}{U_{V}}}$$
$$U_{V} = \frac{(X_{max} - X_{min})^{2}}{12}$$
$$x_{m} = \frac{X_{max} + X_{min}}{2}$$

Note that U_V is used as an approximation of the sum of squares for *X*. It is the variance of a uniform random variable with boundaries X_{min} and X_{max} and σ_X^2 is the variance of the measurement error of *X*. This

procedure makes the assumption that the variances of the measurement errors are approximately equal so that $\sigma_X^2 = \sigma_Y^2$. Note that σ_X^2 is estimated from the data using

$$\hat{\sigma}_X^2 = \frac{1}{2N} \sum_{i=1}^N (x_{1i} - x_{2i})^2$$

The power is computed as

Power =
$$T(-\infty < t_B)$$

where T(x) is the CDF of a Student's t random variable with df = N - 2.

If the sample size is required, this power formula can be used in a binary search.

Power of the Test that the A = 0

The power of this test is computed as follows. First, compute t_A using

$$t_A = \frac{A}{c_A \sigma_X} \sqrt{\frac{N}{2}} - \left| t_{\alpha/2,df} \right|$$

where

$$c_A = \sqrt{1 + \frac{x_m^2}{U_V}}$$

The power is computed as

Power = T(
$$-\infty < t_A$$
)

where T(x) is the CDF of a Student's t random variable with df = N - 2.

If the sample size is required, this power formula can be used in a binary search.

Example 1 – Finding the Sample Size

Clinicians wish to conduct a method comparison study and analyze it with Deming Regression. They want to estimate the number of samples that must be measured when the significance level is 0.05, the power is 0.9, A is 0.2, B is 1.02 to 1.06, min X is 3, max X is 6, and CV is 0.02.

Setup

If the procedure window is not already open, use the PASS Home window to open it. The parameters for this example are listed below and are stored in the **Example 1** settings file. To load these settings to the procedure window, click **Open Example Settings File** in the Help Center or File menu.

Design Tab	
Solve For	.Sample Size
Based On	Both
Power	.0.90
Alpha	.0.05
A (Intercept)	.0.2
B (Slope)	.1.02 to 1.06 by 0.01
Minimum X	.3
Maximum X	.6
Variation Input Type	.CV (Coefficient of Variation)
CV (Coefficient of Variation)	.0.02

Output

Click the Calculate button to perform the calculations and generate the following output.

Numeric Reports

Numeric Results	
Solve For:	Sample Size Based on the Minimum of the Intercept and Slope Powers
Regression Model:	Y = A + BX
Intercept Hypotheses:	H0: A = 0 vs. H1: A \neq 0
Slope Hypotheses:	H0: B = 1 vs. H1: B \neq 1

Power		Sample	Regres Coeffic	Regression Coefficients		X Method Statistics					
Intercept	Slope	N	Intercept	Slope	Min	Мах	Mean	σ	CV	Alpha	
1.00000	0.90030	286	0.2	1.02	3	6	4.5	0.09	0.02	0.05	
0.99697	0.90201	129	0.2	1.03	3	6	4.5	0.09	0.02	0.05	
0.94233	0.90018	73	0.2	1.04	3	6	4.5	0.09	0.02	0.05	
0.90181	0.96059	62	0.2	1.05	3	6	4.5	0.09	0.02	0.05	
0.90181	0.99325	62	0.2	1.06	3	6	4.5	0.09	0.02	0.05	

Power (Intercept) Power (Slope) Power N	The power of the t-test of the intercept regression coefficient. The power of the t-test of the slope regression coefficient. The probability of rejecting a false null hypothesis when the alternative hypothesis is true. The number of samples that are measured by both methods. Two measurements are made using method X and another two measurements are made with method Y on each sample. This results in a total of four measurements per sample.
Intercept	The value of the Y-intercept (A) regression coefficient in the Deming regression model: Y = A + BX. This is the value assumed by the alternative hypothesis. Under the null hypotheses, this value is zero.
Slope	The value of the slope (B) regression coefficient in the Deming regression model: Y = A + BX. This is the value assumed by the alternative hypothesis. Under the null hypotheses, this value is one.
Min X	The minimum measurement value possible for method X. Hence the range of X is [Min, Max]. Note that Min Y = Min X.
Max X	The maximum measurement value possible for method X. Hence the range of X is [Min, Max]. Note that Max Y = Max X.
Mean X	The population mean of the measurements made using method X. Note that Mean X = $(Min X + Max X) / 2$. This procedure makes the simplifying assumption that Mean Y = Mean X.
σ	The population standard deviation of the difference in the duplicate measurements taken on each sample. This procedure makes the simplifying assumption that $\sigma x = \sigma y = \sigma$.
CV	The coefficient of variation of the measurements made by method X. $CV = \sigma x / (Mean X)$. This procedure makes the simplifying assumption that $CVx = CVy = CV$.
Alpha	The probability of rejecting a true null hypothesis.

Summary Statements

A Deming regression (new method Y versus existing method X with both Y and X measured with error) design will be used to test whether the intercept (A) is 0 (H0: A = 0 versus H1: A \neq 0) and whether the slope (B) is 1 (H0: B = 1 versus H1: B \neq 1). The comparisons will be made using Deming regression t-tests of the two coefficients (A and B), with each test having a Type I error rate (α) of 0.05. Duplicate measurements will be made for each method (Y and X) on each sample for a total of four measurements per sample. For both methods, the coefficient of variation is assumed to be 0.02, the standard deviation is assumed to be 0.09, the minimum is assumed to be 3, the maximum is assumed to be 6, and the mean is assumed to be 4.5. To detect an intercept (A) of 0.2 and a slope (B) of 1.02, with at least 90% power for each test, the number of needed subjects will be 286 (the corresponding intercept-test power is 1 and the corresponding slope-test power is 0.9003).

Dropout-Inflated Sample Size

Dropout Rate	Sample Size	Dropout- Inflated Enrollment Sample Size N'	Expected Number of Dropouts D	
20%	286	358	72	
20%	129	162	33	
20%	73	92	19	
20%	62	78	16	
20%	62	78	16	
Dropout Rate	The percentage of sub and for whom no res	jects (or items) that a ponse data will be co	re expected to be llected (i.e., will b	lost at random during the course of the study e treated as "missing"). Abbreviated as DR.
Ν	The evaluable sample are enrolled in the st	size at which power i udv. the design will a	s computed. If N s	subjects are evaluated out of the N' subjects that power.
N'	The total number of su	bjects that should be	enrolled in the stu	udy in order to obtain N evaluable subjects,

based on the assumed dropout rate. After solving for N, N' is calculated by inflating N using the formula N' = N / (1 - DR), with N' always rounded up. (See Julious, S.A. (2010) pages 52-53, or Chow, S.C., Shao, J., Wang, H., and Lokhnygina, Y. (2018) pages 32-33.)

The expected number of dropouts. D = N' - N.

Dropout Summary Statements

Anticipating a 20% dropout rate, 358 subjects should be enrolled to obtain a final sample size of 286 subjects.

References

D

Linnet, Kristian. 1999. 'Necessary Sample Size for Method Comparison Studies Based on Regression Analysis.' Clinical Chemistry, Vol. 45:6, Pages 882-894.

Linnet, Kristian. 1990. 'Estimation of the Linear Relationship Between the Measurements of Two Methods with Proportional Errors.' Statistics in Medicine, Vol. 9, Pages 1463-1473.

These reports show the values of each of the parameters, one scenario per row.

Plots Section



This plot shows the relationship between the slope and sample size.

Example 2 – Validation using Linnet (1999)

Linnet (1999) pages 886-887 presents the following example which we will use to validate this procedure. In this example, the significance level is 0.05, the power is 0.9, A is 0.35, B is 1.058333, min X is 3, max X is 6, and CV is 0.02. The required sample size is 36.

Setup

If the procedure window is not already open, use the PASS Home window to open it. The parameters for this example are listed below and are stored in the **Example 2** settings file. To load these settings to the procedure window, click **Open Example Settings File** in the Help Center or File menu.

Design Tab	
Solve For	Sample Size
Based On	Both
Power	0.90
Alpha	0.05
A (Intercept)	0.35
B (Slope)	1.058333
Minimum X	3
Maximum X	6
Variation Input Type	CV (Coefficient of Variation)
CV (Coefficient of Variation)	0.02

Output

Click the Calculate button to perform the calculations and generate the following output.

Numeric Results

Solve For:Sample Size Based on the Minimum of the Intercept and Slope PowersRegression Model: $Y = A + BX$ Intercept Hypotheses:H0: $A = 0$ vs. H1: $A \neq 0$ Slope Hypotheses:H0: $B = 1$ vs. H1: $B \neq 1$										
Power		Sample	Regre Coeffic	ssion cients		X Me	thod Sta	tistics		
Intercept	Slope	N Size	Intercept	Slope	Min	Max	Mean	σ	CV	Alpha
0.9884	0.90473	36	0.35	1.05833	3	6	4.5	0.09	0.02	0.05

PASS has also calculated the sample size as 36 which validate this procedure.

Note that the power of the intercept test is 0.9884 which is larger than the required 0.9 when N = 36.